



**广 微 测**  
Gmicro Testing

**GUANGDONG DETECTION CENTER OF MICROBIOLOGY**

**REPORT FOR ANALYSIS**

**Report №.**

2021FM21158R01E

**Name of Sample**

ZUTTO HEPA FILTER

**Applicant**



**Test Type**

Entrustment Test



**Address: Building 66, No.100 , Xianlie Middle Road, Guangzhou, China**

**Postcode: 510070**

**Tel: +86 20 87137666**

**Fax : +86 20 87137668**

**Website : www.gddcm.com**



GUANGDONG DETECTION CENTER OF MICROBIOLOGY



REPORT FOR ANALYSIS

Report №.:2021FM21158R01E Verification Code: 23987061

Name of Sample	ZUTTO HEPA FILTER	Test Type	Entrustment Test
Applicant	[REDACTED]	Address	[REDACTED]
Sample Source	Submitted for Testing by the Applicant	Sample Quantity	1 pc
Spec and Lot № of Sample	001	State and Characteristic	Solid
Sample Received Date	2021-10-12	Test Completion Date	2021-11-01
Test Standard and Method	Refer to YY 0469-2011 Appendix B		
Item Tested	Bacterial filtration efficiency		
Test Conclusion	The test data of the sample(s) is attached to the page(s) of this report.		
Remarks	Supplier: [REDACTED] (provided by the applicant)		

Issue Date: 2021-11-15  
(Official Seal)



Editor:

Zeng Yirang

Verifier:

[Signature]

Approver:

Xie Xiaobao



GUANGDONG DETECTION CENTER OF MICROBIOLOGY

ANALYSIS AND TEST RESULT

Report No.: 2021FM21158R01E

No.	Test Item	Test Organisms	Result	Unit
1	Bacterial filtration efficiency	<i>Staphylococcus aureus</i> ATCC 6538	99.96	%
			99.96	
			99.91	

Photo of the sample



(Blank below)

分析  
测专用



## Notice Items

1. The Test report is invalid if not affixed with Authorized Stamp of Test and Paging Seal.
2. The Test report is invalid without signature of verifier and approver.
3. The Test report is invalid if being supplemented, deleted or altered.
4. Without prior written permission, the report cannot be reproduced, except in full.
5. Unless otherwise stated, the results shown in this test report refer only to the sample(s) submitted.
6. Any dispute of the report must be raised to the testing body within 15 days after the report is received, exceeding which the dispute will not be accepted.
7. For the tested sample(s) submitted by the applicant, the sample information in the test report is provided by the applicant and the laboratory is not responsible for its authenticity.
8. The test report without the certification mark (CMA), which involves items that have not obtained the certification, is only used for scientific research, teaching or internal quality control.

